



Programação dos Reagentes CEPA

ABBOTT VP®

ALT-GPT Cepa

Cinético UV-U/L - 37°C (IFCC)

NAME INDEX – TEST NAME	14 - TGP	ASSAY FACTOR	13000
TEMPERATURE	37°C	REAG. BLANK	YES
FILTER CODE	43 (340 – 380)	END POINT	NO
UNITS	07 (U/L)	BGN PRT REVOLUTION	3
DILUTION 1:	26	PRT REVOLUTION	1
REV. TIME	2	INITIAL REAGENT ABS	0,9
AUX. DISP	NO	UP LIMIT	NO
FRR	YES	MAX ABS. LIMIT	1
REACTION UP	NO	SUBSTRATE DEPLETION	0.5
STANDARDS	NO	REAGENT DEGRATION	30