



Programação dos Reagentes CEPA

AIRONE 200®

ALT-GPT Ceba

Cinético UV-U/L - 37°C (IFCC)

TEST NAME	ALT
DECIMAL	0
MEASURE UNIT	U/L
REACTION TYPE	KINETIC
SAMPLE VOLUME (µL)	15 (Soro e Plasma)
REAGENT VOLUME (mL)	
REAGENT 1	400
REAGENT 2	0
ABSORBANCE RANGE (mABS)	
MINIMUM	900
MAXIMUM	2000
LINEARITY LIMIT	250
CONTAMINATING	NO
REAGENT BLANKING	NO
DIFFERENTIAL	NO
FILTERS (nm)	
FILTER 1	340
FILTER 2	NONE
TIMES (SEC)	
MIX 1	0.00
MIX 2	0.00
INCUBATION 1	40
INCUBATION 2	0
LAG PHASE	20
MEASURE	30
MEASUREMENT TYPE	
CALIBRATE	FACTOR
FACTOR	- 1910
NORMAL RANGE	0 - 41