



## Programação dos Reagentes CEPA

### COBAS MIRA / MIRA PLUS

#### ALT-GPT Cepa

Cinético UV-U/L - 37°C (IFCC)

|                  |                 |                      |                   |
|------------------|-----------------|----------------------|-------------------|
| <b>GENERAL</b>   |                 | <b>CALCULATION</b>   |                   |
| MEASUREMENT MODE | <b>ABSORB</b>   | SAMPLE LIMIT         | <b>NO</b>         |
| REACTION MODE    | <b>R-S</b>      | REACTION DIRECTION   | <b>DECREASE</b>   |
| CALIBRATION MODE | <b>FACTOR</b>   | CHECK                | <b>ON</b>         |
| REAGENT BLANK    | <b>REAG/DIL</b> | CONVERSION FACTOR    | <b>1</b>          |
| CLEANER          | <b>NO</b>       | OFF SET              | <b>0</b>          |
| WAVELENGTH       | <b>340</b>      | TEST RANGE LOW       | <b>0</b>          |
| DECIMAL POSITION | <b>0 U/L</b>    | HIGH                 | <b>500</b>        |
| UNIT             | <b>21 U/L</b>   | NORMAL RANGE LOW     | <b>0</b>          |
|                  |                 | HIGH                 | <b>42</b>         |
| <b>ANALYSIS</b>  |                 | NUMBER OF STEPS      | <b>1</b>          |
| POST DIL FACTOR  | <b>2</b>        | CALCULATION STEP A   | <b>KINSEARCH</b>  |
| CONC. FACTOR     | <b>NO</b>       | READING FIRST        | <b>5</b>          |
| SAMPLE CYCLE     | <b>1</b>        | LAST                 | <b>11</b>         |
| VOLUME           | <b>12 µL</b>    | REACTION LIMIT       | <b>NO</b>         |
| DILUTION NAME    | <b>H2O</b>      | <b>CALIBRATION</b>   |                   |
| VOLUME           | <b>20 µL</b>    | CALIBRATION INTERVAL | <b>ON REQUEST</b> |
| REAGENT CYCLE    | <b>1</b>        | TIME                 | <b>NO</b>         |
| VOLUME           | <b>200 µL</b>   | BLANK RANGE LOW      | <b>-</b>          |
|                  |                 | HIGH                 | <b>-</b>          |
|                  |                 | REAGENTE RANGE LOW   | <b>NO</b>         |
|                  |                 | REAGENT RANGE HIGH   | <b>NO</b>         |
|                  |                 | FACTOR               | <b>5776</b>       |