



Programação dos Reagentes CEPA

AIRONE 200®

ASLO Ceba
Turbidimetria

TEST NAME	ASLO
DECIMAL	0
MEASURE UNIT	UI/mL
REACTION TYPE	END POINT
SAMPLE VOLUME (µL)	4
REAGENT VOLUME (µL)	
REAGENT 1	400
REAGENT 2	0
ABSORBANCE RANGE (mABS)	
MINIMUM	0
MAXIMUM	2.000
LINEARITY LIMIT	800
CONTAMINATING	NO
REAGENT BLANKING	YES
DIFFERENTIAL	NO
FILTERS (nm)	
FILTER 1	546
FILTER 2	NONE
TIMES (SEC)	
MIX 1	0.20
MIX 2	0.00
INCUBATION 1	180
INCUBATION 2	0
LAG PHASE	2
MEASURE	1
MEASUREMENT TYPE	
CALIBRATE	STANDARD*
FACTOR	