



Programação dos Reagentes CEPA

COBAS MIRA / MIRA PLUS

ASLO Cepa
Turbidimetria

R1-Diluyente
R2-Látex

GENERAL		CALCULATION	
MEASUREMENT MODE	ABSORB	SAMPLE LIMIT	NO
REACTION MODE	R-S-SR1	REACTION DIRECTION	INCREASE
CALIBRATION MODE	CALIBRATOR	CHECK	ON
REAGENT BLANK	REAG/DIL	CONVERSION FACTOR	1
CLEANER	NO	OFF SET	0
WAVELENGTH	550	TEST RANGE LOW	0
DECIMAL POSITION	0 IU/mL	HIGH	800
UNIT	27 IU/mL	NORMAL RANGE LOW	0
ANALYSIS		HIGH	150
		NUMBER OF STEPS	1
POST DIL FACTOR	2	CALCULATION STEP A	ENDPOINT
CONC. FACTOR	NO	READING FIRST	1
SAMPLE CYCLE	1	LAST	6
VOLUME	3	REACTION LIMIT	NO
DILUTION NAME	H2O	CALIBRATION	
VOLUME	20 µL	CALIBRATION INTERVAL	ON REQUEST
REAGENT CYCLE	1	TIME	NO
VOLUME	270 µL	BLANK RANGE LOW	NO
STAR R1 CYCLE	2	HIGH	NO
VOLUME	30 µL	REAGENTE RANGE LOW	NO
DILUTION NAME	H2O	REAGENT RANGE HIGH	NO
VOLUME	40 µL		
STANDARD POS	*		
STD 1	250		

*Inserido pelo usuário