



Programação dos Reagentes CEPA

AIRONE 200®

FR Cepa
Turbidimetria

TEST NAME	FR
DECIMAL	1
MEASURE UNIT	UI/mL
REACTION TYPE	END POINT
SAMPLE VOLUME (µL)	5
REAGENT VOLUME (µL)	
REAGENT 1	400
REAGENT 2	100
ABSORBANCE RANGE (mABS)	
MINIMUM	0
MAXIMUM	2.000
LINEARITY LIMIT	160
CONTAMINANTING	YES
REAGENT BLANKING	NO
DIFFERENTIAL	NO
FILTERS (nm)	
FILTER 1	546
FILTER 2	NONE
TIMES (SEC)	
MIX 1	1.0
MIX 2	2.0
INCUBATION 1	5
INCUBATION 2	150
LAG PHASE	5
MEASURE	1
MEASUREMENT TYPE	CALIBRATION CURVE
FACTOR	0.0
NORMAL RANGE	0.0 – 30.0