



## Programação dos Reagentes CEPA

### COBAS MIRA / MIRA PLUS

#### LDH Cepa

Piruvato – Cinético UV

|                  |                 |                      |                   |
|------------------|-----------------|----------------------|-------------------|
| <b>GENERAL</b>   |                 | <b>CALCULATION</b>   |                   |
| MEASUREMENT MODE | <b>ABSORB</b>   | SAMPLE LIMIT         | <b>NO</b>         |
| REACTION MODE    | <b>R-S</b>      | REACTION DIRECTION   | <b>DECREASE</b>   |
| CALIBRATION MODE | <b>FACTOR</b>   | CHECK                | <b>ON</b>         |
| REAGENT BLANK    | <b>REAG/DIL</b> | CONVERSION FACTOR    | <b>1</b>          |
| CLEANER          | <b>NO</b>       | OFF SET              | <b>0</b>          |
| WAVELENGTH       | <b>340</b>      | TEST RANGE LOW       | <b>0</b>          |
| DECIMAL POSITION | <b>0 U/L</b>    | HIGH                 | <b>1500</b>       |
| UNIT             | <b>21 U/L</b>   | NORMAL RANGE LOW     | <b>207</b>        |
|                  |                 | HIGH                 | <b>414</b>        |
| <b>ANALYSIS</b>  |                 | NUMBER OF STEPS      | <b>1</b>          |
| POST DIL FACTOR  | <b>5</b>        | CALCULATION STEP A   | <b>KINSEARCH</b>  |
| CONC. FACTOR     | <b>NO</b>       | READING FIRST        | <b>1</b>          |
| SAMPLE CYCLE     | <b>1</b>        | LAST                 | <b>7</b>          |
| VOLUME           | <b>5 µL</b>     | REACTION LIMIT       | <b>NO</b>         |
| DILUTION NAME    | <b>H2O</b>      | <b>CALIBRATION</b>   |                   |
| VOLUME           | <b>10 µL</b>    | CALIBRATION INTERVAL | <b>ON REQUEST</b> |
| REAGENT CYCLE    | <b>1</b>        | TIME                 | <b>NO</b>         |
| VOLUME           | <b>150 µL</b>   | BLANK RANGE LOW      | <b>-0,02</b>      |
|                  |                 | HIGH                 | <b>0,02</b>       |
|                  |                 | REAGENTE RANGE LOW   | <b>NO</b>         |
|                  |                 | REAGENT RANGE HIGH   | <b>NO</b>         |
|                  |                 | FACTOR               | <b>7540</b>       |